

Certificate of Analysis

Product: **recombinant Protein A-PolyHRP20, #AP20C**

Concentration: **1 mg/ml total conjugate**

Buffer base: **0,1M K-Phosphate buffer pH 6,0 + 50% ultra-pure glycerol**

Stabilizing additives: **HRP stabilizers (proprietary) - 100 & 250 µg/ml**
Ultra pure IgG-free BSA - 15 mg/ml

Preservative/anti-microbial: **5-Bromo-5-Nitro-1,3-Dioxane - 600 ppm**

Lot: **705111**

Storage: **-20°C (-15°C/-25°C). Do not store below -25°C!**

Expiration date: **05/2014**

Quality Control Assurance Release Procedure: **ELISA use test with hIgG-biotin as analyte at different conditions**

Result: **POSITIVE (COMPLIES)** ~~**NEGATIVE (DOES NOT COMPLY)**~~

Date: **24.04.2013**

This document has been electronically produced and is valid without a signature