

Certificate of Analysis

Product: **recombinant Protein A-PolyHRP20, #AP20D50**

Concentration: 50 µg/ml pre-diluted **total conjugate** in liquid formulation stabilized for storage at +4°C (format 1/20)

Buffer base: Streptavidin-PolyHRP Stabilizer **#SA1**

Stabilizing additives: HRP stabilizers (proprietary) - 100 & 250 µg/ml
Ultra pure IgG-free BSA - 15 mg/ml

Preservative/anti-microbial: 5-Bromo-5-Nitro-1,3-Dioxane - 600 ppm

Lot: 798358

Storage: -20°C (-15°C/-25°C). Do not store below -25°C!

Expiration date: 12/2007+ (this product is stable up to 6 years and probably longer when stored at **-20°C** in neutral/borosilicate type I glass or low oxygen permeability **PET/PETG/PC** or regular **PP/PE-HD** bottles/vials)

Quality Control Assurance Release Procedure: ELISA use test with hIgG-biotin as analyte at different conditions

Result:

**POSITIVE
(COMPLIES)**

~~**NEGATIVE
(DOES NOT COMPLY)**~~

Date:

27.11.2006

This certificate is a computer printout and has therefore not been signed by hand