

Certificate of Analysis

Product: **recombinant Protein A-PolyHRP40, #AP40C**

Concentration: **1 mg/ml total conjugate**

Buffer base: **0,1M K-Phosphate buffer pH 6,0 + 50% ultra-pure glycerol**

Stabilizing additives: **HRP stabilizers (proprietary) - 100 & 250 µg/ml**
Ultra pure IgG-free BSA - 15 mg/ml

Preservative/anti-microbial: **5-Bromo-5-Nitro-1,3-Dioxane - 600 ppm**

Lot: **977842**

Storage: **-20°C (-15°C/-25°C). Do not store below -25°C!**

Expiration date: **12/2007+ (this product is stable up to 6 years and probably longer when stored at -20°C in neutral/borosilicate type I glass or low oxygen permeability PET/PETG/PC or regular PP/PE-HD bottles/vials)**

Quality Control Assurance
Release Procedure: **ELISA use test with hIgG-biotin as analyte at different conditions**

Result:

**POSITIVE
(COMPLIES)**

~~**NEGATIVE
(DOES NOT COMPLY)**~~

Date:

27.11.2006

This certificate is a computer printout and has therefore not been signed by hand